Maintenance test report

Costumer: Speciallæge Edgar Lauritzen
Address: Frederikssundsvej 14A, 1. th.
Postal code: 2400 **City:** København NV
Country: Danmark

Date: 14-11-2017
Contact: Edgar Lauritzen
Phone: 35817573

RMA No.: 0046 **Case opened:** 14-11-17

Brand: Ukendt **Model:** Ukendt **Serial No.:** Ukendt

Start-up test performed: X **Result:**
De-Ion filter replaced: **Optical components inspected/cleaned:** X **Electrical connections inspected:** X
Water filter replaced: **Optical components aligned:** **Safety functions tested:** X
Air filter replaced: **Laser effect measured:** **Mechanical safety functions tested:** X
Cooling liquid(s) refilled: **Laser effect adjusted:**

Unit meets all manufactor specifications: **Unit approved:** X
Comments:

Kontrol enhed på væg er forsynet med hovedafbryder, der virker. Selve lysgiver enheden består af totalt 6 stk. Dialight Lumidrives af typen LSP9-R-005 a' 1.03 watt pr. stk. Der er 9 LED pr. enhed. Samlet forbrug af hver enhed er målt til 9.2 W, hvilket viser at der ikke er målt degenerering på nogle af de 6 enheder. Se yderligere!

N.B.:Checked if necessary or done!

Repairs / service performed which is not included in Service agreement:

Replaced parts: **Hours:** **Km.:** **Transp:** **Date:**

Final comments:

Komplet test af alle funktioner og timere udført. Test af lysdiodernes strømforbrug.
Eftersyn af elektriske forbindelser. Enheden er godkendt.

Technician: BS
Case closed: 14-11-17